

S-5019-03

UCHL-1
CD3 > control
CD20
CD79

Fax: 709-777-6061

Attn: Dr. Ejeckam

M7193 anti-CD3, clone PC3/188a

07.07.98 detex survey: L2 = 1/10 (2-.5), EnVis = L+ = no signal at 1/10-1/25
PI-78

- tonsil
- HIER: MW-PC, 3 mins std., TRF (S1699) / 1/10 - 1/100, 30 min. / L2 (K0675), EnVis. (K1491), L+ (K0690), 30 mins. ea. reagent / DAB (D1300), 10 mins.
- surprisingly weak (no) signal, espec. w/ L+ and EnVis. @ 1/10.

07.14.98 HIER survey: HpH (S3307) required for HIER PI-78B

- tonsil
- HIER: 1) MW-PC, 3 mins std vs. MW-AR, 2x5 mins. each 20 mins. RT
2) TRF vs. HpH
- CD3 @ 1/10, 30 mins. f.b. L2 or EnVis. 30 mins. ea. reagent
- CD3 @ 1/25, 30 mins. f.b. L+, 30 mins. ea. reagent
- HpH (S3307) >> TRF (S1699),

thy ^{control} musSharon 506 648 6516
ST John's, New Brunswick.