

Dr. Donald Cook

From: Good, Nancy [NGood@mtsina.on.ca]
Sent: Wednesday, September 21, 2005 5:13 PM
To: Dr. Donald Cook
Subject: ER-PR reports

Hi Dr. Cook,

I spoke to Dr. Mullen and he will read the slides on the weekend and have reports for you early next week.

bye for now,
Nancy

Ejection to point man

Sept 22, 2005 meeting with Trish, Dr. Williams, Kam Lang, Sullivan, Pridmore

- recommendation external CAP (Need this)
- have frustrated overwhelmed individuals
- have metabolic technological
- don't understand theory, work as a group
- need to have good line of communication, ~~no~~ need one technical director
- ~~no~~ technical competencies written by tech - Section Head - need a pathologist
- Need chief tech for immunohistology, - dedicated staff,

↓
Need manuals signed off by Clinical Chief

- Need better fixation = standardized fixator
- Need pathology assistants
- outside labs - need to ~~p~~ outside labs need to advise of their fixation.
- ∴ HCCSJ needs to advise other hospitals of fixation

consider disclaimer on reports for referral labs

- Controls - Need to run non-ve and -ve controls
- non-ve -ve controls for specificity
- non-ve controls to ensure stain is working
- if internal control are not working, revisit controls

} run on every patient block

- Validation - no validation forms on
- Trish will learn SOP
- No record books, no documentation
- Need to start writing manual, need computer skills
- Need to validate reagent kits, document expiry dates

9/22/2005

- Need Textbooks, internet access, arrange lines of communication with Halifax

- No pipette calibrated

- Need Manuals - approval by Medical Director

= Need storage for IF - Need -70°C

= No DATA Sets

- need internet

= Pathologists : communicates

: supply Journals

: tech need to know why ~~the~~ thing are done immuno

= Purchase : Histo columns

: insert on microtome

: alignment lesson for immuno

: Need for small lesson lessons

⊗ documentate for equipment maintenance

Competency testing for technologists

Sends someone for MSH every year
external CAP

Her-2-Neu - need a control

∴ Humidity temperature in the room

There are 4-D steps in these area

Need a Succession plan