

Dr. Donald Cook

From: Dr. Donald Cook
Sent: Wednesday, September 21, 2005 9:32 AM
To: 'NGood@mtsina.on.ca'
Subject: ER-PR cases

Hi Nancy

Could you advise me on the status of the ER and PR slides, and when we can expect some reports.

Regards,
Dr Cook

Spoke to Trust, Dr Bw Carls Sept 21, 2005 11:00 AM

Start with
CAP program testing

Trust

Tech are overwhelmed

Need to validate lots

Need to document pH

- Humidity in temperature & freeze
- Need basic building block
- Pipettes need to be calibrated

Dabo
system
worked fine

at least a
year to get
the thing
up and
properly

- No calibration of Dabo system
- No assurance things were done the same way
- No standardization of Dabo system
- No validation of system
- don't know when to look to trouble shoot
- Need Kim or May to be in charge of some immuno assays
- NSH - National Society of histopathology - provide good lecture

② Need

① need specialized immunotech
to ^{do} validation procedure

③ policy manual, as written by tech themselves

④ - Need to be trained how to use controls

⑤ Keep control down on slides

⑥ Fixation - need standardized fixative, cut some thickness, need some processing

⑦ Need protocol manual for fixation

⑧ Things are being done, but no document

⑨ Need to be trained to do trouble shoot

⑩ - Need to get involved with CAP program

⑪ Tech don't have the skill

⑫ Need training at computer

⑬ Need PA's

⑭ Need spec sheets

⑮ Need extend QA program

⑯ Need to send to well organized, with some feedback

9/21/2005