

RAININ Original Certificate of Conformity

Rainin Instrument, LLC, as an ISO 9001 accredited company, hereby certifies that the enclosed pipette has been inspected and validated according to the RAININ Quality Assurance System. This system utilizes statistical process control in assuring the quality of design, production and control procedures for each instrument.

ISO 9001:2000
ACCREDITED

Although the long-term reliability of your pipette depends on its original design and the quality of the production process, a final gravimetric test assures that your pipette will perform within its specifications. Each pipette is inspected by accredited technicians, under strictly defined and monitored conditions:

- Air temperature ($21.5^{\circ}\text{C} \pm 1.0^{\circ}\text{C}$). Humidity (45 - 75%).
- Balances are regularly serviced and checked with weights traceable to a national standard.

Accuracy of each instrument is individually checked by a series of gravimetric weighings and performance validated prior to shipment. This is indicated by a PASS status on the Performance Check Report.

Specifications rely on the quality and consistency for the entire pipetting system; they are guaranteed when using RAININ-manufactured tips. Satisfactory performance also requires the user to be acquainted with the instructions detailed in the Instruction Manual.

RAININ

PO Box 2160, Oakland, CA 94621-9949 • 510-564-1600
a METTLER TOLEDO Company

9920-219 Rev J



BUSINESS REPLY MAIL


FIRST CLASS MAIL PERMIT NO. 412 OAKLAND, CA

POSTAGE WILL BE PAID BY ADDRESSEE

Rainin Instrument, LLC

PO Box 2160
Oakland, CA 94621-9949

NO POSTAGE
NECESSARY
IF MAILED
IN THE
UNITED STATES



Rainin Test Report

Gravimetric Data

Identification

Model: L-1000
 Serial No.: K0601691A
 Rainin Tip No.: RT-L1000
 Useful Volume Range: 100µl - 1000µl

Test Conditions

Technician: LL
 Balance No.: AT261
 Balance Sensitivity (g): 0.00001
 Correction Factor (µl/mg): 1.0031
 Correction for Evaporation (µl): 0.000
 Basis of adjustment: Ex (to deliver)
 Temperature: 21.5 ± 1.0°C
 Barometric Pressure: 101kPa
 Relative Humidity: 50%

Volume Setting (µl)	100	1000
Weighings (mg)	100.3	995.9
	101.3	996.9
	101.0	996.6
	101.1	997.1
Mean (mg)	100.9	996.6
Mean (µl)	101.2	999.7
Systematic Error (µl) (accuracy)	101.2	999.7
Random Error (µl) (precision)	0.4	0.5

Test Results

Low Volume Setting (µl):	100	Pass
High Volume Setting (µl):	1000	Pass

Declaration of conformity to the ISO 8655 Standard

Type-testing of Rainin pipettes has been performed by qualified technicians in accordance with the ISO 8655-6 standard. Results are maintained on file.

Gravimetric data in mg is converted to volumetric results in µl for distilled water and evaporation is accounted for if applicable.

The PASS status indicates that this pipette performs within Rainin's specifications. Therefore the performance is also within the maximum permissible errors allowed by ISO 8655-2.

Limited Warranty and Limitation of Liability

a.) Seller warrants the merchandise to conform to specifications. Under no circumstances shall Seller be responsible for alleged nonconformities with respect to any merchandise which has been used for purposes or in any manner for which it was not intended, or any merchandise which has been customized or modified without Seller's prior written consent, or damaged or misused. As Buyer's exclusive remedy in the event of breach of warranty, Seller shall repair or replace, as its option, any nonconforming merchandise or parts thereof for a period of one (1) year after delivery. All claims must be made in writing to the Seller. Any claims not made within the period specified above shall be deemed waived and released.

b.) THE PROVISIONS OF THE FOREGOING WARRANTIES ARE IN LIEU OF ANY OTHER WARRANTY, WHETHER EXPRESS OR IMPLIED, WRITTEN OR ORAL (INCLUDING ANY WARRANTY OF MERCHANTABILITY, FITNESS FOR A PARTICULAR PURPOSE, TITLE OR NON-INFRINGEMENT). SELLER'S LIABILITY ARISING OUT OF THE MANUFACTURE, SALE OR SUPPLYING OF A PRODUCT OR ITS USE OR ITS DISPOSITION, WHETHER BASED UPON WARRANTY, CONTRACT, TORT OR OTHERWISE, SHALL NOT FOR ANY REASONS EXCEED THE AGGREGATE PURCHASE PRICE PAID BY BUYER FOR SUCH PRODUCT. IN NO EVENT SHALL SELLER BE LIABLE TO BUYER OR ANY OTHER PERSON OR ENTITY FOR SPECIAL, INCIDENTAL, CONSEQUENTIAL OR EXEMPLARY DAMAGES (INCLUDING, BUT NOT LIMITED TO, DAMAGES FOR LOSS OF PROFITS, LOSS OF DATA OR LOSS OF USE) ARISING OUT OF THE MANUFACTURE, SALE, SUPPLY, USE, MARKETING, RESALE OR OPERATION OF THE MERCHANDISE, EVEN IF SELLER HAS BEEN ADVISED OF THE POSSIBILITY OF SUCH DAMAGES OR LOSSES.

Model: L-1000
 Serial No.: K0601691A

WARRANTY REGISTRATION

To ensure registration of warranty, please complete and return this postage-paid card. **PLEASE PRINT CLEARLY**

This postage-paid warranty return card is for use in the U.S. only. Outside the U.S., place the card in an envelope and send it to your Rainin distributor.

LAST NAME										FIRST NAME										M.I.			
COMPANY/ INSTITUTION																							
DEPARTMENT								BUILDING								ROOM				MAIL STOP			
MAILING ADDRESS																							
CITY												STATE						ZIP					
PHONE						FAX						E-MAIL											

DATE OF PURCHASE

DO YOU WANT TO RECEIVE NEW PRODUCT INFORMATION BY E-MAIL? ☐ YES ☐ NO
 DO YOU WANT TO RECEIVE NEW PRODUCT INFORMATION BY FAX? ☐ YES ☐ NO

BUSINESS TYPE

(SELECT MOST APPROPRIATE)

- ☐ UNIVERSITY/
 MED SCHOOL
☐ PHARMACEUTICAL
☐ INDUSTRIAL

- ☐ CLINICAL CHEMISTRY/
 DIAGNOSTICS
☐ BIOTECH
☐ GOVERNMENT

- ☐ FOOD/BEVERAGE
☐ ENVIRONMENTAL
☐ FORENSICS

TYPE: ☐ OTHER: