

**ER/PR Test Results with Dako Autostainer in Use
Statistical Performance**

Date	Total Tests²	Total Pos.	% Pos.	Total Neg.²	% Neg.	False Neg.³	% False Neg.
May-Dec 31, 1997	137	80	58%	57	42%	16	28%
1998	147	71	48%	76	52%	51	67%
1999	360	234	65%	126	35%	71	56%
2000	370	200	54%	170	46%	49	29%
2001 ¹	374	231	62%	143	38%	61	43%
2002	344	197	57%	147	43%	71	48%
	1732	1013	58%	719	42%	319	44%

Ejeckam Committee Intervention

Date	Total Tests²	Total Pos.	% Pos.	Total Neg.²	% Neg.	False Neg.³	% False Neg.
2003	373	284	76%	89	24%	39	44%
Jan-April 1, 2004	109	93	85%	16	15%	11	69%
	482	377	78%	105	22%	50	48%
Grand Totals	2214	1390	63%	824	37%	369	45%

Test Results while Ventana Autostainer in Use

Date	Total Tests²	Total Pos.	% Pos.	Total Neg.²	% Neg.	False Neg.³	% False Neg.
April-Dec 31, 2004	381	340	89%	41	11%	1	.02%
Jan-July 31, 2005	114	95	83%	19	17%	2	10%
Totals	495	435	88%	60	12%	3	.05%

(Prepared by Dr. Charles Hutton for Ches Crosbie in relation to May 10, 2007 and August 3, 2007 Answers to Interrogatories in action 2006 01T 2966 CP)

¹ From May 1997 to December 2000 the cut off point for positivity was 30%. This was reduced to 10% in 2001.

² Exhibit P-1852, answer 3

³ Exhibit P-1853, answer 7